

The handiest glue gas analyser from ecom for inspections and shortterm measurements on small combustion systems

Tested in accordance with EN 50379-2 and 1st BlmSchV



COMPACT FLUE GAS ANALYSIS

Made in Germany



Reliable

Precise measurement results thanks to sensor calibration in the climate chamber



Efficient

Switch on - measure - manage data - done!





THE HANDIEST ECOM SOLUTION

Checking and adjusting gas and oil heating systems



- Freehand measurement thanks to magnets on the back for attachment to metallic surfaces
- CO sensor overload protection with free purging without measurement interruption
- Long battery operation time with up to 150 measurements without charging thanks to integrated Li-lon battery
- Up to 300 measurement data records storable

• = Basis EC • = Optional EC

O₂

CO

NO

Technische Daten			√ Standard • Op	tion
Measured values	Range	Resolution	Accuracy *= Higher value prevails	
Maximum amoun	t gas sensors			3
02	021 %	0,1 vol. %	± 0,3 vol. %	√
CO (H ₂ -comp.)	02.500 ppm (10.000 ppm)	1 ppm	± 20 ppm / 5 % of the measured value*	√
NO	05.000 ppm	1 ppm	± 5 ppm / 5 % of the measured value*	
Other measured variables	Range	Resolution	Accuracy	
T-Gas	0500 °C	0,1 °C	± 2 °C or 1,5 % of the measured value*	√
T-Air	099 °C	0,1 °C	±1°C	√
Pressure △P	± 100 hPa	0,01 hPa	± 0,5 hPa or 1 % of the measured value*	√

Technical data	
Calculation values	Range
CO ₂	0CO _{2 max}
Combustion efficiency (ETA)	0120 %
Excess air (Lambda)	>1
Losses qA	0100 %
Dew point	x° C
mg/m³	x mg/m³
mg/kWh	x mg/kWh
O ₂ reference	x % O ₂

Transport bag

Spacious textile bag for analyser and tools.



ecom-P

Infrared thermal printer.



CO-multi-hole probe

For measuring the CO concentration.



Annular gap multi-hole probe

For measuring the $\rm O_2$ concentration in the annular gap.



Soot pump set

To determine the soot dot number.



ecom-xRE

Readout head for digital burner controllers..



ecom-UNO

Pocket-sized differential pressure gauge.



e.CLOUD by ecom

Digital measurement data and customer management.

